



# STYLEVIEW™

by ergotron®

## Cleaning Recommendations

You can use most hospital-grade, diluted, non-abrasive, non-corrosive cleaning solutions on all plastic and metal surfaces, whether painted, polished or unfinished. This includes quaternary ammonia compounds, ammonia enzyme cleaners, hydrogen peroxide, bleach or alcohol solutions.

### Helpful tips:

- Be sure to test any cleaning solution on a small, hidden area before use.
- Pen and permanent and dry erase markers should be removed with 91% isopropyl alcohol and a soft cloth.
- Iodine stains can be removed with commonly used cleaners and a soft cloth.

### For your safety and to protect your medical cart:

- Never use steel wool or other abrasive materials that will damage the surface finish.
- Do not use strong solvents such as trichloroethylene and acetone as they can damage the surface finish.
- To avoid the risk of electric shock, do not expose electrical components to water, cleaning solutions or other potentially corrosive liquids or substances.
- Do not immerse the cart or cart components in liquid or allow liquids to flow into the cart. Wipe all cleaners off the surface immediately using a damp cloth and dry the surface after cleaning.
- Do not use flammable cleaners on cart surfaces due to the proximity to electrical power.



Be sure to consult your hospital's infection control administrator or epidemiologist for all cleaning procedures and processes.

